

calibration & metrology



About us




Overview

- Europe's largest owner managed metrology laboratory
- Established in 1976
- Manufacturer independent
- Accredited since 1997:
D-K-15019-01-00,
DIN EN ISO/IEC 17025:2018






About us



esz AG in figures


- 25.000 satisfied customers
- 200.000 calibrations per year
- Over 200 accredited quantities
- 45 years of experience
- 7 locations in Europe
- 243 employees
- 2 mio Euro annual investment





Teamwork – Our employees

- 243 employees at 7 locations
- Colleagues from 20 nations
- Continuous growth (2009: 63 > 2021: 201)
- Above-average period of employment
- High level of expertise





Our locations




Here you find esz AG

- Eichenau (Headquarter)
- Nuremberg
- Mannheim
- Steinfurt
- Berlin
- Vienna (Austria)
- Budapest (Hungary)





Business Units



Calibration

- Accredited:
DAKKS D-K-15019-01-00
DIN EN ISO/IEC 17025:2018
- Factory calibration /
ISO-compliant calibration
- Customized
- CEPS:
Calibration Expert Partner System



Service

- On-site
- Repair
- Safety inspection / DGUV-V3
- Consulting
- Training
- Pick-up and delivery service
- Tool Control Systems
- Equipment leasing
cooperation with



Software

- **calibration expert:** Calibration management
- **asset expert:** Test & Measurement equipment management





Industries



Customers & Markets



Medical Technology



Telecommunications



Automotive



Automotive inspection bodies



Environment



Aviation



Mechanical Engineering



Electronics



IT / Software



Transportation / Logistics



Energy



Accreditations



Deutsche Akkreditierungsstelle GmbH

Berehene gemäß § 8 Absatz 1 AkkStelleG i.V.m. § 1 Absatz 1 AkkStelleGBV
Unterzeichnerin der Multilateralen Abkommen
von EA, ILAC und IAF zur gegenseitigen Anerkennung

Akkreditierung



Die Deutsche Akkreditierungsstelle GmbH bestätigt hiermit, dass das Kalibrierlaboratorium
esz AG calibration & metrology
Max-Planck-Straße 16, 82223 Eichenau

die Kompetenz nach DIN EN ISO/IEC 17025:2018 besitzt, Kalibrierungen in folgenden Bereichen
durchzuführen:

Elektrische Messgrößen
Gleichstrom- und Niederfrequenzmessgrößen
Zeit und Frequenz

Hochfrequenz- und Strahlungsmessgrößen
Hochfrequenzmessgrößen
Optische Messgrößen

Thermodynamische Messgrößen
Temperaturmessgrößen
Feuchtemessgrößen

Mechanische Messgrößen
Druck
Kraft
Drehmoment
Messgeräte im Kraftfahrtwesen (MIK)
Durchflussmessgrößen

Dimensionelle Messgrößen
Länge
Winkel

Die einzelnen Messgrößen sind in der Anlage enthalten.

Die Akkreditierungsurkunde gilt nur in Verbindung mit dem Bescheid vom 24.11.2020 mit der
Akkreditierungsnr. D-K-15019-01. Sie besteht aus diesem Deckblatt, der Rückseite des
Deckblatts und der folgenden Anlage mit insgesamt 81 Seiten.

Registriernummer der Urkunde: D-K-15019-01-00

Berlin, 24.11.2020

Im Auftrag Dr. Helga Manke
Abteilungsleiterin

Die Urkunde samt Urkundenanlage gibt den Stand zum Zeitpunkt des Ausstellungsdatums wieder. Der jeweils aktuelle Stand
Geltungsbereiches der Akkreditierung ist die Datenbank akkreditierter Stellen der Deutschen Akkreditierungsstelle GmbH (DAkkS) zu
entnehmen. <https://www.dakkis.de/content/datenbank-akkreditierter-stellen>

Urkundenanlage auf der Rückseite

With over 200 accredited quantities, esz AG is one of the leading laboratories in Europe. The scope of accreditations is growing every year.



Electrics



Optics



Acoustics




Mechanics



Flow



Pressure



Temperature



Humidity



Force



Mass



Torque




Acceleration



Medical Technology



KFZ test benches



Illuminance



Calibration




DAkkS- and ISO-Calibration

- DAkkS accredited calibration
- ISO calibration according to the requirements of DIN EN ISO/IEC 17025:2018
- Traceability, specifications and quality characteristics, conformity assessment, procedures and requirements are identical
- Fast and flexible order processing in own laboratories
- Calibration certificate available digitally as pdf document (full version) and/or as paper document (short version)



Calibration Certificate



Digital Calibration Certificate

- Full evaluation with detailed measurement data
- Download with QR code or certificate data via
 - esz-calibration certificate portal: www.esz-ag.de/kalibrierscheindownload
 - Test equipment management software asset expert

Hardcopy Calibration Certificate

- On customer request
- Content:
 - Calibration object and method
 - Ambient and measurement conditions
 - Place of calibration
 - Measurement uncertainties, traceability chain and due date
 - Evaluation, whether given results meet specified requirements or not




Calibration Certificate



In detail

■ Identical service level for accredited or non-accredited calibrations

- Traceability, specifications and quality characteristics, conformity assessment, procedures as well as requirements
- Difference: Cover sheet of the calibration certificate



■ Live monitoring of the ambient conditions

- Real-time readings of temperature, humidity, air pressure
- Accurate evaluation during calibration in the esz laboratory



Calibration Certificate

Calibration Mark: 405364-01-D-K-15019-2021-01

1. Device Under Test

Manufacturer: Keysight
Model: 34401A
Model type: Digital Multimeter
AC
DC
Type: benchtop device
Application(s): n.a.
Inspection equipment No.: n.a.
Inventory No.: n.a.

All stated measurement and test results relate only to the item mentioned above.

2. Calibration Procedure

- Direct measurement procedure using fixed standards or a variable AC/DC-source according to QMH III.1
- Frequency calibration at an external synchronized generator according to esz QMH VIII.1.2

Calibration procedure revised: .../.../..., approved 2015-09-24 by Wilhelm Sandmeier

Calibration equipment and standards:

Standard	Manufacturer Model	Device	Traceable to	Cal. no.	Last cal.	Next cal.
034195	Fluke 579B	Calibrator	D-K-15019-01-00	034195-14	2021-01	2022-01
090800	Precision Test Systems GPS10eR	Frequency Standard	GPS-10eR	090800-20	2019-01	2020-01
116398	esi 4-Wire-Cu Sheet	Short	D-K-15019-01-00	116398-91	2019-01	2020-01

Auxiliary equipment and devices:

Aux. devices	Manufacturer Model	Device	Traceable to	Cal. no.	Last cal.	Next cal.
063278	Agilent 33220A	Function Generator	D-K-15019-01-00	063278-08	2020-06	2021-06

Calibration Mark: 405364-01-D-K-15019-2021-01

DC-Resistance Zero Offset Tests, 4-Wire

Range	Nominal value	Measured value	Specification	Deviation	%TOL	MU	Rem.	Diagram
10.0	0.00000 0	0.000 0	+ 0.00	0.0000 0	20 %	6.059×10^{-2} 0		●
1.0	0.000000 0	0.000001 00	+ 0.00	0.000001 00	10 %	6.059×10^{-2} 0		●
10.0	0.000000 0	0.000001 00	+ 0.00	0.000001 00	10 %	6.059×10^{-2} 0		●
100.0	0.000000 0	0.00001 00	+ 0.00	0.00001 00	10 %	6.059×10^{-2} 0		●
1.00	0.000000 00	0.000000 00	+ 0.00	0.000000 00	0 %	6.059×10^{-2} 0		●
10.00	0.000000 00	0.000000 00	+ 0.00	0.000000 00	0 %	6.059×10^{-2} 0		●
100.00	0.000000 00	0.000000 00	+ 0.00	0.000000 00	0 %	6.059×10^{-2} 0		●
1.000	0.000000 00	0.000000 00	+ 0.00	0.000000 00	0 %	6.059×10^{-2} 0		●
10.000	0.000000 00	0.000000 00	+ 0.00	0.000000 00	0 %	6.059×10^{-2} 0		●
100.000	0.000000 00	0.000000 00	+ 0.00	0.000000 00	0 %	6.059×10^{-2} 0		●

Calibration Mark: 405364-01-D-K-15019-2021-01

DC-Voltage

Range	Nominal value	Measured value	Specification	Deviation	%TOL	MU	Rem.	Diagram
10.0V	100.0000 0V	106.000 0V	+ 0.200 % + 0.000 NPS	-0.0000 0	0 %	12×10^{-2}		●
100.0V	1000.0000 0V	1019.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	12×10^{-2}		●
1V	0.000000 0	0.000000 0	+ 0.200 % + 0.000 NPS	-0.5000 0	5 %	14×10^{-2}		●
10.0V	-1.00000 0V	-1.00000 0V	+ 0.200 % + 0.000 NPS	-0.0000 0	0 %	14×10^{-2}		●
100.0V	-100.0000 0V	-106.000 0V	+ 0.200 % + 0.000 NPS	-6.0000 0	10 %	12×10^{-2}		●
10.00V	-0.000000 00	-0.000000 00	+ 0.200 % + 0.000 NPS	-0.5000 0	5 %	13×10^{-2}		●
100.00V	-1000.0000 0V	-1019.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	13×10^{-2}		●
10V	-0.000000 00	-0.000000 00	+ 0.200 % + 0.000 NPS	-0.5000 0	5 %	13×10^{-2}		●
100V	-10000.0000 0V	-10190.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	13×10^{-2}		●
1000V	-100000.0000 0V	-101900.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	13×10^{-2}		●
10000V	-1000000.0000 0V	-1019000.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	13×10^{-2}		●
100000V	-10000000.0000 0V	-10190000.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	13×10^{-2}		●
1000000V	-100000000.0000 0V	-101900000.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	13×10^{-2}		●
10000000V	-1000000000.0000 0V	-1019000000.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	13×10^{-2}		●
100000000V	-10000000000.0000 0V	-10190000000.000 0V	+ 0.200 % + 0.000 NPS	-1.0000 0	10 %	13×10^{-2}		●

Calibration Mark: 405364-01-D-K-15019-2021-01

esz AG calibration & metrology

Max-Planck-Str. 14 | D-87773 Füssen | +49-811-98805-0 | info@esz-ag.de | www.esz-ag.de

Report Revision 0354-01 (replacing all previous 1.221-01-08)
Digital copies are generated after visually and digitally signed.

5 / 11

In detail

Detailed traceability information

- Representation of the uninterrupted traceability chain of the calibration equipment and standards
- Guaranteed metrological traceability to the SI system of units or to the underlying metrological standards

Graphic chart for each specified measurement step


- Representation as a line diagram
- Compliance with the tolerances of the device under test relative to the nominal value at a glance




Conformity statement

Detailed risk evaluation

- Verification of the specific risk of the conformity statement for each measurement step for accredited calibrations (DAkkS) or factory calibrations (ISO).
- Six cases can occur:
 - **Without remarks:** >95 % probability of being within the specification (ILAC G8:09/2019 Fig.5, Pass)
 - **!:** Warning threshold exceeded, >95 % probability of being within the specification
 - **?:** with a probability of >50% to 95% of being within the specification, taking into account the measurement uncertainty (MU) (ILAC G8:09/2019 Fig.5, Conditional Pass)
 - **?!:** above the warning threshold exceeded but with a probability of >50% to 95% of being within the specification, taking into account the MU
 - **X?:** with a probability of >50% to 95% of being outside the specification (ILAC G8:09/2019 Fig.5, Conditional Fail), taking into account the MU
 - **X:** >95% probability of being outside the specification (ILAC G8:09/2019 Fig.5, Fail)





Conformity statement

■ Decision Rules

- Selected “ILAC evaluation“ defines the decision rule and indication for each measuring step
 - FAIL:** Decision rule according to DIN EN ISO 14253-1:2018-5.2 (on customer request). Measurement results according to ILAC-G8:09/2019 Fig. 5, Pass are assessed as conforming to the specifications. Cases according to ILAC-G8:09/2019 Fig. 5, Conditional Pass, Conditional Fail and Fail are not considered to comply with the specifications.
 - PASS:** Standard decision rule according to the QM system of esz AG. Measurement results according to ILAC-G8:09/2019 Fig. 5, Pass und Conditional Pass are assessed as conforming to the specifications. Cases according to ILAC-G8:09/2019 Fig. 5, Conditional Fail und Fail are not considered to comply with the specifications.
 - None:** Decision rule according to ILAC evaluation „PASS“ but without risking, i.e. question marks do not show during calibartion and in the calibration certificate.



Detailed reporting

■ Comprehensive data and terms in all esz calibration certificates

- Deviation: Difference between measured value and nominal value
 - Rem.: Indication per measurement (ILAC evaluation)
 - Intervention/warning threshold: 80 % of specification per default, adaptable to customer requirements
 - Statement of Conformity: statement whether given results meet specified requirements or not (successful conformity assessments per measurement step)
 - MU: Measurement uncertainty, determined according to EA-4/02 M:2013
 - Specification: Definition of the properties of a device by the user:
 - **%rdg:** Percent of measured value – **D:** Digit (least significant digit of the display) – **FS:** Full Scale (range end value) – **ppm:** parts per million (10^{-6})
 - **ppb:** parts per billion (10^{-9}) – **rdg:** reading (display)




Additional Services

calibration & metrology

accredited by the Deutsche Akkreditierungsstelle (DAkkS)
Member of the Deutschen Kalibrierdienst


Calibration Certificate No. 405364-01

Object	Digital Multimeter Keyence 3401A	Passed
Serial No.	123456789	
Customer	esz Marketing Max-Planck-Str. 16 82223 Eichenau	
esz ID	405364	
Test Equipment No.	n.a.	
Inventory No.	P0123456	
Customer Ref. No.	CD2021	
Order No.		
Traceability	This calibration certificate documents the traceability to the International System of Units (SI) Systeme International d'Unités). Quality management system principles and calibration procedures comply with DIN EN ISO/IEC 17025. The calibration fulfills the requirements according to DIN EN ISO 9001 H. or other comparable quality management guidelines.	
Date of Calibration	2021-01-26	
Next Calibration	2022-01	The calibration laboratory does not make any suggestion on the calibration interval. The definition and compliance to intervals for recalibrations lie within the user's responsibility.
Approved by	Andreas Böck	
Laboratory Manager	Person in Charge	
  		
<small>Report Revision 4.0/04-01 (Impresario) at pre-revision 1, 2021-01-18 Digital copies were generated automatically and are digitally signed.</small>		
<small>esz AD calibration & metrology Max-Planck-Str. 16, D-82223 Eichenau +49 89 8007-2 info@esz.de www.esz.de</small>		
<small>Report Revision 4.0/04-01 (Impresario) at pre-revision 1, 2021-01-18 Digital copies were generated automatically and are digitally signed.</small>		

calibration & metrology

accredited by the Deutsche Akkreditierungsstelle (DAkkS)
Member of the Deutschen Kalibrierdienst

Calibration Certificate No. 405364-02

Object	Digital Multimeter Keyence 3401A	Passed
Serial No.	123456789	
Customer	esz Marketing Max-Planck-Str. 16 82223 Eichenau	
esz ID	405364	
Test Equipment No.	n.a.	
Inventory No.	P0123456	
Customer Ref. No.	CD2021	
Order No.		
Traceability	This calibration certificate documents the traceability to the International System of Units (SI) Systeme International d'Unités). Quality management system principles and calibration procedures comply with DIN EN ISO/IEC 17025. The calibration fulfills the requirements according to DIN EN ISO 9001 H. or other comparable quality management guidelines.	
Date of Calibration	2021-01-27	
Next Calibration	2022-01	The calibration laboratory does not make any suggestion on the calibration interval. The definition and compliance to intervals for recalibrations lie within the user's responsibility.
Approved by	Andreas Böck	on 2021-01-27
Laboratory Manager	Person in Charge	
  		
<small>Report Revision 4.0/04-01 (Impresario) at pre-revision 1, 2021-01-18 Digital copies were generated automatically and are digitally signed.</small>		
<small>esz AD calibration & metrology Max-Planck-Str. 16, D-82223 Eichenau +49 89 8007-2 info@esz.de www.esz.de</small>		
<small>Report Revision 4.0/04-01 (Impresario) at pre-revision 1, 2021-01-18 Digital copies were generated automatically and are digitally signed.</small>		

Calibration certificates with customer logo

- Available on request for accredited calibrations (DAkkS) and factory calibrations (ISO)
- Factory calibration certificates also available without esz-Logo on request





Additional Services

Driftbericht

Erstellungsdatum 24.11.2020

esz-ID: 103109 Seriennummer: 174810342
 Modell: HBM 1-P3TCP / 200 bar abs. Top Prüfmittelnummer: 3490
 Class:
 Gegenstand: Drucksensor
 Kunde: RAPA Rausch & Pausch GmbH

Inventarnummer: n.a.

Die Spalten enthalten die ermittelten Abweichungen zum am besten zuordbaren Kalibrierwert der jeweiligen Kalibrierung.

Eintrag Erklärung

- keine Abweichung im Rahmen der Stellenauflösung des Messwertes
- n.a. keine Drift aufgrund fehlender Überschreitung der Historiendaten berechenbar
- no drift keine Drift im Rahmen der Stellenauflösung des Messwertes feststellbar
- max. betragsmäßig größter Wert aller Driften der ermittelten Historie
- Drift mittl. verzeichnenrichtiger Mittelwert aller Driften
- Bem. Drift Bem. unter Berücksichtigung der maximalen Drift liegen die zu erwartenden Werte der nächsten Folgekalibrierung außerhalb der Toleranzen
- ! Bemerkung zur Risikoabschätzung unter Berücksichtigung der maximalen Drift liegen die zu erwartenden Werte der nächsten Folgekalibrierung außerhalb der Toleranz

Druck, Absolut-

Bereich	Einstellung	Kalibrierwert	103109-02	103109-03	max. Drift	mittl. Drift	Bem.
200 bar	0,973 mV/V	40,0 bar	0,2 bar	0,1 bar	0,10 bar	0,1 bar	
200 bar	0,9742 mV/V	80,0 bar	0,2 bar	0,1 bar	0,10 bar	-0,1 bar	
200 bar	1,1344 mV/V	120,0 bar	0,2 bar	0,2 bar	no drift	no drift	
200 bar	1,5942 mV/V	160,0 bar	0,1 bar	0,1 bar	no drift	no drift	
200 bar	1,5953 mV/V	200,0 bar	-0,1 bar	0,1 bar	0,20 bar	0,2 bar	

esz AG calibration & metrology
 Max-Planck-Str. 15 D-82223 Eichenau +49-8141-88887-0 info@esz-ag.de www.esz-ag.de

Report Revision 200828 56.2 (ersetzt alle Vervorherige)


Seite 1/1

Drift report

- Available on customer request
- Appendix to the digital calibration certificate
- Information on deviations of the measured values compared to previous calibrations
- Risk estimation for the upcoming calibration interval
- Line by line representation for each specified measurement step



Software Solutions




asset expert

- Web-based software solution
- Cross-company management of test equipment management
- The architecture of the software allows any access, anytime, anywhere.
- Independent of the hardware configuration
- Design and "look and feel" individually customizable.
- Hierarchy levels, user structures and rights management can be set according to user requirements.
- Demo version <https://www.esz-ag.de/software/pruefmittelmanagement.html>



Software Solutions



asset expert

The screenshot displays the asset expert software interface. At the top, there is a header bar with the esz logo and the text "asset expert". Below the header, the main content area is divided into several sections:


- PRÜFMITTELDETAILS**: A table showing detailed information about a measurement tool.

INFORMATIONEN ZUM PRÜFMITTEL	
esz-ID	000001
Fälligkeitsdatum	2021-05-19
Prüfmittelnr.	KL13082
Inventarnr.	LAB10
Seriennr.	US28032894
Hersteller	Hewlett Packard
Gegenstand	Digitalmultimeter
Typ/Modell	3458A
Überwacht	<input checked="" type="checkbox"/>
Kalibrierscheinnr.	000001-15
Kalibrierdatum	2020-05-19
Intervall	12
Status	OK
Zustand	kalibriert
- ARTIKELDETAILS**: A table showing details for an article.

Artikelnr.	HP-D3458A
Firma	esz AG calibration & metrology
Kundennummer	60000
Benutzername	p.fleischmann@esz-ag.de
Bemerkung	
Abteilung	Ei
Garantie	18.03.2008
Leistungsumfang	DKD
Bezugsnormal?	
Lieferant	T.O.P
Preis	7900 EUR
Lieferdatum	18.03.2007
Standort	LAB7
- KALIBRIERHISTORIE**: A table showing a history of calibrations.

KALIBRIERSCHEINNR.	AUFRAGNSNR.	KALIBRIERDATUM	INTERVALL	PRÜFER	ERGEBNIS
000001-15	BK130297	2020-05-19	12	Sanktjohanser, Andreas	Erfolgreich
000001-14	BK106351	2019-05-31	12	Brzezinski, Przemyslaw	Erfolgreich
000001-13	BKO85532	2018-05-28	12	Sanktjohanser, Andreas	Erfolgreich
000001-12	BKO67816	2017-05-09	12	Sanktjohanser, Andreas	Erfolgreich
000001-11	BKO53822	2016-05-20	12	Brzezinski, Przemyslaw	Erfolgreich
- PRÜFMITTELLISTE**: A table showing a list of measurement tools.


ESZ-ID	#	FÄLLIGKEITSDATUM	PRÜFMITTELNR.	SERIENNR.	HERSTELLER	GEGENSTAND	TYPE/MODEL	KALIBRIERSCHEINNR.	INTERVALL	STATUS	ABTEILUNG	LEISTUNGSUMFANG	STANDORT
000001	1	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-15	12	OK	Ei	DKD	LAB10
000002	2	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-14	12	OK	Ei	DKD	LAB10
000003	3	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-13	12	OK	Ei	DKD	LAB10
000004	4	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-12	12	OK	Ei	DKD	LAB10
000005	5	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-11	12	OK	Ei	DKD	LAB10
000006	6	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-10	12	OK	Ei	DKD	LAB10
000007	7	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-09	12	OK	Ei	DKD	LAB10
000008	8	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-08	12	OK	Ei	DKD	LAB10
000009	9	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-07	12	OK	Ei	DKD	LAB10
000010	10	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-06	12	OK	Ei	DKD	LAB10
000011	11	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-05	12	OK	Ei	DKD	LAB10
000012	12	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-04	12	OK	Ei	DKD	LAB10
000013	13	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-03	12	OK	Ei	DKD	LAB10
000014	14	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-02	12	OK	Ei	DKD	LAB10
000015	15	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-01	12	OK	Ei	DKD	LAB10
000016	16	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-00	12	OK	Ei	DKD	LAB10
000017	17	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-099	12	OK	Ei	DKD	LAB10
000018	18	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-100	12	OK	Ei	DKD	LAB10
000019	19	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-101	12	OK	Ei	DKD	LAB10
000020	20	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-102	12	OK	Ei	DKD	LAB10
000021	21	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-103	12	OK	Ei	DKD	LAB10
000022	22	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-104	12	OK	Ei	DKD	LAB10
000023	23	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-105	12	OK	Ei	DKD	LAB10
000024	24	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-106	12	OK	Ei	DKD	LAB10
000025	25	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-107	12	OK	Ei	DKD	LAB10
000026	26	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-108	12	OK	Ei	DKD	LAB10
000027	27	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-109	12	OK	Ei	DKD	LAB10
000028	28	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-110	12	OK	Ei	DKD	LAB10
000029	29	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-111	12	OK	Ei	DKD	LAB10
000030	30	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-112	12	OK	Ei	DKD	LAB10
000031	31	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-113	12	OK	Ei	DKD	LAB10
000032	32	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-114	12	OK	Ei	DKD	LAB10
000033	33	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-115	12	OK	Ei	DKD	LAB10
000034	34	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-116	12	OK	Ei	DKD	LAB10
000035	35	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-117	12	OK	Ei	DKD	LAB10
000036	36	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-118	12	OK	Ei	DKD	LAB10
000037	37	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-119	12	OK	Ei	DKD	LAB10
000038	38	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-120	12	OK	Ei	DKD	LAB10
000039	39	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-121	12	OK	Ei	DKD	LAB10
000040	40	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-122	12	OK	Ei	DKD	LAB10
000041	41	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-123	12	OK	Ei	DKD	LAB10
000042	42	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-124	12	OK	Ei	DKD	LAB10
000043	43	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-125	12	OK	Ei	DKD	LAB10
000044	44	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-126	12	OK	Ei	DKD	LAB10
000045	45	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-127	12	OK	Ei	DKD	LAB10
000046	46	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-128	12	OK	Ei	DKD	LAB10
000047	47	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-129	12	OK	Ei	DKD	LAB10
000048	48	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-130	12	OK	Ei	DKD	LAB10
000049	49	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-131	12	OK	Ei	DKD	LAB10
000050	50	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-132	12	OK	Ei	DKD	LAB10
000051	51	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-133	12	OK	Ei	DKD	LAB10
000052	52	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-134	12	OK	Ei	DKD	LAB10
000053	53	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-135	12	OK	Ei	DKD	LAB10
000054	54	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-136	12	OK	Ei	DKD	LAB10
000055	55	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-137	12	OK	Ei	DKD	LAB10
000056	56	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-138	12	OK	Ei	DKD	LAB10
000057	57	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-139	12	OK	Ei	DKD	LAB10
000058	58	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-140	12	OK	Ei	DKD	LAB10
000059	59	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-141	12	OK	Ei	DKD	LAB10
000060	60	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-142	12	OK	Ei	DKD	LAB10
000061	61	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-143	12	OK	Ei	DKD	LAB10
000062	62	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-144	12	OK	Ei	DKD	LAB10
000063	63	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-145	12	OK	Ei	DKD	LAB10
000064	64	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-146	12	OK	Ei	DKD	LAB10
000065	65	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-147	12	OK	Ei	DKD	LAB10
000066	66	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-148	12	OK	Ei	DKD	LAB10
000067	67	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-149	12	OK	Ei	DKD	LAB10
000068	68	2021-05-19	KL13082	US28032894	Hewlett Packard	Digitalmultimeter	3458A	000001-150	12	OK	Ei</td		



Equipment service and repair

- Manufacturer-independent repair hub
- Extensive service catalog
- Archive with over 20,000 service documents/manuals
- Service cooperation with well-known manufacturers
- Transparent cost overview for repairs
- Acceleration of repairs thanks to spare parts flat rate
- Faster order placement thanks to design certificate
- esz adjustment is only charged if 100% possible





Safety test / DGUV-V3

- Manufacturer-independent safety testing of electrical measuring equipment according to DGUV V3 as part of the calibration order or as an individual order.
- Safety tests of electrical equipment in three sections according to the DGUV V3 test protocol:
 - visual inspection
 - functional test
 - measurement check
- Legally compliant test protocol
- Marking of the device with test sticker





Consulting

- Feasibility check for metrological projects
- Calibration procedure descriptions and measurement uncertainty budgets
- Implementation in calibration expert
- Training of esz departments and external partners
- Development of calibration procedures and measuring stations for esz AG
- Accreditation sharing and authorization network coordination for external organisations





Service



Trainings

- esz competence training
- Seminars
- Trainings
- Employee trainings
- Symposium: Munich Calibration Day





Service



Logistik Expert / On-site

- Professional pick-up and transport service in Germany, Austria, Hungary, Poland and Benelux
- Short downtimes
- Trained specialist personnel
- Own equipment-specific transport packaging
- Full value insurance
- Daily delivery routes
- Cost transparency for on-site service
- Onsite calibration at the customer's location:
 - at the customer's premises
 - in own laboratory containers

[Click here](#) for the detailed list of delivery routes






Storage and transport solutions

Best for tools, test equipment and other items

- Development and production of storage systems for any desired tool or device
- Individually milled foam boards (Shadow Boards)
- Laser engraving of tools, mats, cases
- Equipped TCS according to your specifications
- Shadow boards available in an ESD version on request.



The plus in safety in workshops and on the road

- Efficient, mobile and individual solutions for all industries
- Uncompromising security thanks to optimal storage solutions
- Avoidance of Foreign Object Damage e.g. during aircraft maintenance



Safety needs **confidence**. We make
confidence **measurable**.
For today's **decisions**
and **products**
for the **future**.